
Midmark M11 - v1.0.5
Total Cycles: 548

Name: RITTER2
Sterilizer ID: V2313228
11/15/2021 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.1 F 0.0
2:00 133.1 F 1.2
4:00 136.0 F 3.7
6:00 165.4 F 7.9
8:00 209.2 F 12.4
10:00 224.5 F 8.0
12:00 231.8 F 13.2
14:00 236.8 F 10.5
16:00 234.0 F 8.0
18:00 247.6 F 14.8
20:00 258.9 F 21.1
22:00 269.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 30.5
1:00 272.0 F 30.5
2:00 272.1 F 30.5
3:00 272.0 F 30.4
4:00 271.3 F 30.5

Min 271.2 F 30.4
Max 272.4 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 227.7 F 0.0
10:00 225.8 F 0.0
15:00 221.0 F 0.0
20:00 219.6 F 0.0
25:00 222.8 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:49
CYCLE COMPLETE

11/15/2021 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____