
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001243
14:34 01/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:49
STERILIZING 3:54
136.0 C 220kPa 3:54
136.4 C 226kPa 4:24
136.4 C 225kPa 4:54
136.4 C 224kPa 5:24
136.4 C 224kPa 5:54
136.4 C 222kPa 6:24
136.3 C 223kPa 6:54
136.2 C 223kPa 7:24
136.2 C 223kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 23:14

Digital Signature #
1D35CF62D437AE8C
