
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006991
9:35 27/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:57
STERILIZING 3:58
136.4 C 223kPa 3:58
136.3 C 224kPa 4:28
136.2 C 224kPa 4:58
136.5 C 223kPa 5:28
136.1 C 220kPa 5:58
136.3 C 220kPa 6:28
136.4 C 223kPa 6:58
136.2 C 222kPa 7:28
136.2 C 222kPa 7:28
VENTING 7:28
AIR DRYING 8:20
CYCLE COMPLETE 8:33

Digital Signature #
2672628934E43E01
