
Midmark M11 - v1.0.5
Total Cycles: 1664

Name: RITTER2
Sterilizer ID: V2313228
10/26/2022 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.7 F 0.0
2:00 184.7 F 3.7
4:00 192.3 F 8.5
6:00 225.3 F 8.8
8:00 230.8 F 10.2
10:00 231.7 F 7.9
12:00 240.9 F 11.6
14:00 244.8 F 13.6
16:00 258.0 F 20.5
18:00 269.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.2 F 29.8
2:00 271.1 F 29.3
3:00 272.5 F 29.9
4:00 271.3 F 29.3

Min 271.1 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 221.2 F 0.0
10:00 237.2 F 0.0
15:00 235.1 F 0.0
20:00 228.5 F 0.0
25:00 224.7 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

10/26/2022 04:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____