
Midmark M11 - v1.0.5
Total Cycles: 1909

Name: RITTER2
Sterilizer ID: V2313228
01/10/2023 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.0 F 0.0
2:00 80.4 F 0.0
4:00 114.2 F 0.0
6:00 141.4 F 0.0
8:00 151.6 F 0.1
10:00 180.4 F 0.8
12:00 202.6 F 4.3
14:00 216.9 F 7.9
16:00 229.3 F 11.9
18:00 229.4 F 7.6
20:00 241.2 F 12.7
22:00 239.6 F 10.7
24:00 236.7 F 9.2
26:00 250.8 F 15.8
28:00 262.0 F 22.3
30:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.2
2:00 271.6 F 28.9
3:00 272.3 F 29.2
4:00 271.5 F 28.7

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 213.4 F 0.0
10:00 224.5 F 0.0
15:00 235.4 F 0.0
20:00 242.2 F 0.0
25:00 230.6 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:30
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

01/10/2023 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____