

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000822  
16:30 28/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:53  
PRESSURIZING 6:33  
STERILIZING 8:12  
136.5 C 224kPa 8:12  
136.4 C 221kPa 8:42  
136.4 C 221kPa 9:12  
136.1 C 221kPa 9:42  
136.3 C 222kPa 10:12  
136.2 C 220kPa 10:42  
136.4 C 223kPa 11:12  
136.1 C 220kPa 11:42  
136.1 C 220kPa 11:42  
VENTING 11:42  
AIR DRYING 12:44  
CYCLE COMPLETE 35:02  
-----

Digital Signature #  
9FF5C2699259B223  
-----