
Midmark M11 - v1.0.0
Total Cycles: 918

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2021 06:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 77.4 F 0.0
4:00 75.9 F 0.0
6:00 95.0 F 0.1
8:00 131.1 F 0.2
10:00 168.5 F 0.2
12:00 194.1 F 1.3
14:00 209.0 F 4.3
16:00 218.1 F 6.9
18:00 227.2 F 10.0
20:00 237.1 F 10.8
22:00 236.8 F 10.1
24:00 239.2 F 9.0
26:00 245.1 F 12.5
28:00 245.3 F 12.8
30:00 257.6 F 19.4
32:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.7
2:00 272.1 F 28.8
3:00 272.0 F 28.3
4:00 272.5 F 29.0

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 219.4 F 0.0
10:00 220.1 F 0.0
15:00 201.2 F 0.0
20:00 242.1 F 0.0
25:00 228.1 F 0.0
30:00 218.1 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 33:09
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:11:14
CYCLE COMPLETE

06/21/2021 08:14 AM
ACCEPTED: X
REJECTED:
Operator: U01