

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001374  
13:07 01/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:57  
PRESSURIZING 3:59  
STERILIZING 5:12  
136.0 C 221kPa 5:12  
136.2 C 223kPa 5:42  
136.2 C 223kPa 6:12  
136.3 C 224kPa 6:42  
136.4 C 223kPa 7:12  
136.1 C 223kPa 7:42  
136.3 C 224kPa 8:12  
136.4 C 225kPa 8:42  
136.4 C 225kPa 8:42  
VENTING 8:42  
AIR DRYING 9:39  
CYCLE COMPLETE 18:15  
-----

Digital Signature #  
FC0699E1228811BE  
-----