
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000506
15:55 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 3:56
STERILIZING 5:24
136.4 C 232kPa 5:24
136.5 C 233kPa 5:54
136.4 C 231kPa 6:24
136.3 C 232kPa 6:54
136.4 C 233kPa 7:24
136.2 C 231kPa 7:54
136.2 C 232kPa 8:24
136.4 C 231kPa 8:54
136.4 C 231kPa 8:54
VENTING 8:54
AIR DRYING 9:53
CYCLE COMPLETE 10:11

Digital Signature #
7DC7145E461B1731
