
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
3.4uS / 2.1ppm
CYCLE NUMBER 000774
11:52 16/02/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:16
PRESSURIZING 5:58
STERILIZING 7:22
136.0 C 220kPa 7:22
136.3 C 219kPa 7:52
136.1 C 221kPa 8:22
135.8 C 221kPa 8:52
136.4 C 225kPa 9:22
135.6 C 220kPa 9:52
136.3 C 224kPa 10:22
136.3 C 224kPa 10:52
136.3 C 224kPa 10:52
VENTING 10:52
AIR DRYING 11:47
CYCLE COMPLETE 41:47

Digital Signature #
FD7406135179F3BD
