
Midmark M11 - v1.0.0
Total Cycles: 3129

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2023 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.7 F 0.0
2:00 195.6 F 1.3
4:00 219.1 F 6.9
6:00 230.5 F 12.7
8:00 232.8 F 9.0
10:00 240.5 F 9.9
12:00 245.7 F 13.4
14:00 245.1 F 12.7
16:00 258.4 F 19.8
18:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.7
2:00 271.3 F 28.2
3:00 272.6 F 28.9
4:00 271.7 F 28.2

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 219.8 F 0.0
10:00 233.4 F 0.0
15:00 228.7 F 0.0
20:00 226.1 F 0.0
25:00 224.0 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/19/2023 12:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____