
Midmark M11 - v1.0.0
Total Cycles: 3877

Name: RITTER 1
Sterilizer ID: V2178198
10/06/2023 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 82.2 F 0.0
4:00 107.8 F 0.0
6:00 129.6 F 1.5
8:00 159.3 F 3.9
10:00 181.4 F 6.8
12:00 205.7 F 10.0
14:00 217.3 F 10.9
16:00 221.2 F 6.3
18:00 232.1 F 10.1
20:00 234.0 F 8.7
22:00 241.5 F 11.6
24:00 237.8 F 9.3
26:00 251.7 F 16.0
28:00 263.2 F 22.7
30:00 271.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.2 F 28.4
2:00 272.9 F 29.0
3:00 271.9 F 28.4
4:00 271.9 F 28.3

Min 271.5 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 192.7 F 0.0
10:00 192.5 F 0.0
15:00 196.7 F 0.0
20:00 223.5 F 0.0
25:00 237.1 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 30:08
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:07:54
CYCLE COMPLETE

10/06/2023 10:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____