

-----  
Midmark M11 - v1.0.0  
Total Cycles: 612

Name: MIDMARK  
Sterilizer ID: V2178198  
01/20/2021 11:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.8 F 0.0  
2:00 166.2 F 2.8  
4:00 186.4 F 7.2  
6:00 207.1 F 12.7  
8:00 221.5 F 6.4  
10:00 235.6 F 11.9  
12:00 234.9 F 8.9  
14:00 238.8 F 9.5  
16:00 247.2 F 14.0  
18:00 260.1 F 21.0  
20:00 271.6 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.4  
1:00 272.7 F 29.0  
2:00 271.7 F 28.8  
3:00 272.2 F 28.6  
4:00 272.6 F 29.2

Min 271.4 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.8 F 0.0  
5:00 215.4 F 0.0  
10:00 239.2 F 0.0  
15:00 231.6 F 0.0  
20:00 220.7 F 0.0  
25:00 218.5 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 20:26  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 00:58:30  
CYCLE COMPLETE

01/20/2021 12:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_