

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2565

Name: RITTER 1  
Sterilizer ID: V2178198  
11/16/2022 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
2:00 188.9 F 3.2  
4:00 201.7 F 7.6  
6:00 212.9 F 12.5  
8:00 225.2 F 7.2  
10:00 237.8 F 12.7  
12:00 235.9 F 9.8  
14:00 238.3 F 8.5  
16:00 248.5 F 14.7  
18:00 261.0 F 21.7  
20:00 272.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.7 F 29.5  
2:00 272.0 F 28.8  
3:00 271.7 F 29.1  
4:00 272.6 F 29.5  
  
Min 271.4 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 232.1 F 0.0  
10:00 236.1 F 0.0  
15:00 229.5 F 0.0  
20:00 228.0 F 0.0  
25:00 230.3 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:45  
CYCLE COMPLETE

11/16/2022 11:11 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_