
Midmark M11 - v1.0.5
Total Cycles: 2928

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 174.6 F 2.8
4:00 186.8 F 6.9
6:00 206.5 F 11.6
8:00 223.4 F 6.1
10:00 230.3 F 11.0
12:00 232.6 F 7.8
14:00 244.8 F 12.4
16:00 244.9 F 12.7
18:00 256.5 F 19.0
20:00 267.1 F 25.6
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.0 F 29.0
2:00 272.5 F 29.2
3:00 272.5 F 29.4
4:00 272.1 F 29.6

Min 271.5 F 28.9
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 236.1 F 0.0
10:00 223.7 F 0.0
15:00 225.0 F 0.0
20:00 216.6 F 0.0
25:00 219.0 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

09/26/2023 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____