
Midmark M11 - v1.0.0
Total Cycles: 2115

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2022 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.3 F 0.0
2:00 193.5 F 3.7
4:00 212.2 F 8.1
6:00 215.0 F 12.8
8:00 224.4 F 7.4
10:00 238.4 F 12.5
12:00 237.9 F 9.9
14:00 235.9 F 8.5
16:00 249.1 F 14.7
18:00 260.9 F 21.4
20:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.2
2:00 272.5 F 28.9
3:00 271.6 F 28.4
4:00 271.1 F 28.4

Min 271.1 F 28.2
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 225.6 F 0.0
10:00 244.1 F 0.0
15:00 228.6 F 0.0
20:00 225.0 F 0.0
25:00 223.6 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

07/11/2022 03:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____