

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000552  
13:27 23/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:54  
STERILIZING 4:04  
136.0 C 228kPa 4:04  
136.3 C 232kPa 4:34  
136.4 C 233kPa 5:04  
136.1 C 228kPa 5:34  
136.4 C 231kPa 6:04  
136.3 C 231kPa 6:34  
136.2 C 231kPa 7:04  
136.2 C 230kPa 7:34  
136.2 C 230kPa 7:34  
VENTING 7:35  
AIR DRYING 8:26  
CYCLE COMPLETE 14:45

-----  
Digital Signature #  
5C92AE13EFDF663F  
-----