
Midmark M11 - v1.0.0
Total Cycles: 3566

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2023 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	70.1 F	0.0
2:00	74.2 F	0.0
4:00	92.2 F	0.0
6:00	114.3 F	0.0
8:00	140.0 F	1.1
10:00	169.7 F	2.7
12:00	185.6 F	5.0
14:00	199.5 F	8.1
16:00	210.5 F	11.5
18:00	216.8 F	4.3
20:00	226.4 F	7.7
22:00	236.2 F	11.5
24:00	232.0 F	7.4
26:00	244.2 F	12.8
28:00	241.4 F	10.8
30:00	252.7 F	16.5
32:00	262.7 F	22.3
34:00	272.6 F	28.8

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	28.1
1:00	272.4 F	28.6
2:00	271.4 F	28.0
3:00	272.5 F	28.8
4:00	271.6 F	28.0

Min 271.4 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	212.1 F	0.0
5:00	201.6 F	0.0
10:00	197.8 F	0.0
15:00	191.7 F	0.0
20:00	191.9 F	0.0
25:00	194.0 F	0.0
30:00	196.7 F	0.0

Drying Complete

FILLING: 0:50
HEATING: 34:25
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:12:40
CYCLE COMPLETE

07/24/2023 09:37 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____