

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001618  
7:45 26/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:56  
STERILIZING 4:04  
136.5 C 223kPa 4:04  
136.3 C 223kPa 4:34  
136.3 C 220kPa 5:04  
136.2 C 221kPa 5:34  
136.3 C 223kPa 6:04  
136.4 C 223kPa 6:34  
136.4 C 223kPa 7:04  
136.4 C 221kPa 7:34  
136.4 C 221kPa 7:34  
VENTING 7:34  
AIR DRYING 8:26  
CYCLE COMPLETE 8:30

-----  
Digital Signature #  
58223797038EA297  
-----