
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001502
10:53 10/14/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 5:02
STERILIZING 6:26
136.4 C 221kPa 6:26
136.2 C 221kPa 6:56
136.3 C 222kPa 7:26
136.3 C 222kPa 7:56
135.8 C 220kPa 8:26
136.3 C 220kPa 8:56
136.3 C 222kPa 9:26
136.1 C 221kPa 9:56
VENTING 9:56
AIR DRYING 11:00
CYCLE COMPLETE 41:00

Digital Signature #
DD5A263FC1A71106
