
Midmark M11 - v1.0.5
Total Cycles: 773

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.3 F 0.0
2:00 159.3 F 2.8
4:00 176.7 F 7.0
6:00 196.2 F 12.3
8:00 222.4 F 7.0
10:00 232.0 F 12.6
12:00 237.3 F 10.2
14:00 235.7 F 8.7
16:00 250.3 F 15.8
18:00 262.3 F 22.8
20:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.1 F 29.3
2:00 271.9 F 29.6
3:00 271.8 F 29.1
4:00 271.9 F 29.7

Min 271.3 F 28.9
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 210.0 F 0.0
10:00 216.8 F 0.0
15:00 208.6 F 0.0
20:00 201.0 F 0.0
25:00 203.5 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

02/08/2022 04:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____