

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001615  
13:49 20/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:31  
STERILIZING 3:32  
136.0 C 220kPa 3:32  
136.4 C 225kPa 4:03  
136.4 C 225kPa 4:33  
136.4 C 224kPa 5:03  
136.4 C 225kPa 5:33  
136.4 C 224kPa 6:03  
136.4 C 226kPa 6:33  
136.4 C 226kPa 7:03  
136.4 C 226kPa 7:03  
VENTING 7:03  
AIR DRYING 7:57  
CYCLE COMPLETE 8:01  
-----

Digital Signature #  
2DE85E63EF0419BC  
-----