
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003106
10:13 08/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:18
STERILIZING 4:32
136.4 C 224kPa 4:32
136.3 C 224kPa 5:02
136.3 C 223kPa 5:32
136.4 C 220kPa 6:02
136.3 C 223kPa 6:32
136.3 C 223kPa 7:02
136.4 C 222kPa 7:32
136.4 C 224kPa 8:02
136.4 C 224kPa 8:02
VENTING 8:03
AIR DRYING 8:58
CYCLE COMPLETE 9:03

Digital Signature #
1BADBE685A0C4A78
