

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2471

Name: RITTER 1  
Sterilizer ID: V2178198  
10/21/2022 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.2 F 0.0  
2:00 164.7 F 3.8  
4:00 200.8 F 7.9  
6:00 211.6 F 12.6  
8:00 215.0 F 6.3  
10:00 228.5 F 11.1  
12:00 230.0 F 7.3  
14:00 243.6 F 13.4  
16:00 242.7 F 11.9  
18:00 255.9 F 18.7  
20:00 267.5 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 29.4  
2:00 271.9 F 28.7  
3:00 271.9 F 29.0  
4:00 272.6 F 29.3  
  
Min 271.4 F 28.4  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.0 F 0.0  
5:00 213.2 F 0.0  
10:00 231.6 F 0.0  
15:00 230.8 F 0.0  
20:00 227.6 F 0.0  
25:00 227.9 F 0.0  
30:00 230.6 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 21:23  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:32  
CYCLE COMPLETE

10/21/2022 02:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_