
Midmark M11 - v1.0.0
Total Cycles: 2333

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2022 07:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 79.2 F 0.0
4:00 103.4 F 0.0
6:00 123.9 F 0.0
8:00 143.1 F 0.2
10:00 171.9 F 3.5
12:00 192.2 F 6.8
14:00 206.5 F 10.0
16:00 226.3 F 11.1
18:00 229.0 F 7.4
20:00 237.0 F 11.8
22:00 235.7 F 8.4
24:00 243.0 F 11.6
26:00 245.1 F 12.6
28:00 256.9 F 18.7
30:00 267.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.1 F 28.5
2:00 272.3 F 29.0
3:00 271.4 F 28.4
4:00 272.1 F 28.6

Min 271.4 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 195.3 F 0.0
10:00 217.4 F 0.0
15:00 218.3 F 0.0
20:00 217.2 F 0.0
25:00 221.4 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

09/19/2022 08:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____