
Midmark M11 - v1.0.5
Total Cycles: 896

Name: RITTER2
Sterilizer ID: V2313228
03/23/2022 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 177.8 F 0.2
4:00 191.6 F 3.1
6:00 214.0 F 8.3
8:00 227.6 F 10.6
10:00 235.1 F 10.6
12:00 234.8 F 8.7
14:00 239.7 F 9.9
16:00 249.1 F 15.0
18:00 261.8 F 22.3
20:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.7
2:00 272.6 F 29.3
3:00 271.5 F 28.7
4:00 271.4 F 29.1

Min 271.1 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 227.0 F 0.0
10:00 224.5 F 0.0
15:00 215.8 F 0.0
20:00 205.0 F 0.0
25:00 218.3 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

03/23/2022 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____