
Midmark M11 - v1.0.0
Total Cycles: 4423

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2024 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.8 F 0.0
2:00 150.3 F 1.4
4:00 148.8 F 4.2
6:00 172.0 F 7.8
8:00 195.1 F 11.8
10:00 214.1 F 4.1
12:00 224.4 F 8.0
14:00 236.8 F 12.2
16:00 234.3 F 8.7
18:00 242.5 F 11.6
20:00 243.9 F 12.2
22:00 255.2 F 18.1
24:00 265.7 F 24.3
26:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.0
2:00 272.4 F 29.4
3:00 271.3 F 28.7
4:00 272.5 F 29.2

Min 271.2 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 221.9 F 0.0
10:00 204.8 F 0.0
15:00 199.9 F 0.0
20:00 206.7 F 0.0
25:00 207.3 F 0.0
30:00 198.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 26:33
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:23
CYCLE COMPLETE

04/03/2024 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____