
Midmark M11 - v1.0.5
Total Cycles: 2711

Name: RITTER2
Sterilizer ID: V2313228
07/31/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 80.2 F 0.0
4:00 127.7 F 0.0
6:00 112.8 F 0.0
8:00 148.9 F 0.0
10:00 177.6 F 0.0
12:00 196.0 F 0.0
14:00 206.7 F 0.0
16:00 211.8 F 0.0
18:00 223.1 F 3.7
20:00 231.6 F 7.0
22:00 240.6 F 10.8
24:00 237.4 F 8.9
26:00 245.5 F 12.4
28:00 245.3 F 12.9
30:00 244.1 F 12.3
32:00 256.8 F 19.0
34:00 267.8 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.3 F 28.6
2:00 272.8 F 29.1
3:00 272.3 F 29.0
4:00 272.0 F 28.4

Min 271.7 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 220.9 F 0.0
10:00 208.9 F 0.0
15:00 196.0 F 0.0
20:00 193.1 F 0.0
25:00 207.5 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 35:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:13:18
CYCLE COMPLETE

07/31/2023 11:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____