
Midmark M11 - v1.0.0
Total Cycles: 2876

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2023 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 198.0 F 4.6
4:00 203.3 F 9.9
6:00 221.3 F 5.8
8:00 232.7 F 10.7
10:00 232.4 F 7.8
12:00 241.3 F 10.9
14:00 246.7 F 13.6
16:00 259.7 F 20.7
18:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.1
2:00 271.8 F 28.4
3:00 271.9 F 28.8
4:00 272.7 F 29.0

Min 271.4 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 227.9 F 0.0
10:00 234.2 F 0.0
15:00 232.9 F 0.0
20:00 229.4 F 0.0
25:00 229.8 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

02/08/2023 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____