
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000036
11:38 19/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:09
STERILIZING 3:16
136.4 C 222kPa 3:16
135.9 C 218kPa 3:46
136.2 C 221kPa 4:16
136.4 C 222kPa 4:46
136.4 C 222kPa 5:16
136.4 C 223kPa 5:46
136.1 C 219kPa 6:16
136.3 C 222kPa 6:46
136.3 C 222kPa 6:46
VENTING 6:46
AIR DRYING 7:31
CYCLE COMPLETE 8:05

Digital Signature #
12D7B603E6625A6A
