
Midmark M11 - v1.0.5
Total Cycles: 2465

Name: RITTER2
Sterilizer ID: V2313228
06/05/2023 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.5 F 0.0
2:00 178.2 F 3.6
4:00 192.3 F 8.2
6:00 216.2 F 11.0
8:00 230.1 F 8.9
10:00 238.5 F 10.4
12:00 242.4 F 12.4
14:00 241.3 F 11.3
16:00 254.1 F 18.0
18:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.9
2:00 272.1 F 29.3
3:00 271.3 F 29.0
4:00 272.1 F 29.8

Min 271.3 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 222.5 F 0.0
10:00 246.2 F 0.0
15:00 240.4 F 0.0
20:00 225.8 F 0.0
25:00 224.2 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

06/05/2023 02:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____