
Midmark M11 - v1.0.5
Total Cycles: 306

Name: RITTER2
Sterilizer ID: V2313228
08/24/2021 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 170.6 F 3.0
4:00 193.3 F 7.8
6:00 214.3 F 13.4
8:00 228.3 F 8.6
10:00 237.4 F 10.5
12:00 243.3 F 12.9
14:00 243.3 F 12.3
16:00 256.7 F 19.2
18:00 268.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.0 F 28.9
2:00 272.4 F 29.2
3:00 272.7 F 29.6
4:00 271.6 F 29.0

Min 271.5 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 226.9 F 0.0
10:00 204.2 F 0.0
15:00 215.6 F 0.0
20:00 180.9 F 0.0
25:00 178.1 F 0.0
30:00 175.3 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

08/24/2021 10:29 AM
ACCEPTED: X
REJECTED:
Operator: U02