

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007393  
11:28 15/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:23  
STERILIZING 4:30  
136.4 C 223kPa 4:30  
136.1 C 222kPa 5:00  
136.4 C 222kPa 5:30  
136.1 C 222kPa 6:00  
136.4 C 224kPa 6:30  
136.4 C 224kPa 7:00  
136.3 C 222kPa 7:30  
136.4 C 223kPa 8:00  
136.4 C 223kPa 8:00  
VENTING 8:00  
AIR DRYING 8:54  
CASSETTE REMOVED  
CYCLE COMPLETE 17:31  
-----

Digital Signature #  
8053D3BCFD0124E0  
-----