
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001554
9:14 01/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:44
STERILIZING 3:50
136.0 C 220kPa 3:50
136.4 C 220kPa 4:20
136.2 C 223kPa 4:50
136.3 C 222kPa 5:20
136.2 C 222kPa 5:50
136.4 C 223kPa 6:20
136.3 C 224kPa 6:50
136.4 C 224kPa 7:20
136.4 C 224kPa 7:20
VENTING 7:20
AIR DRYING 8:12
CYCLE COMPLETE 38:12

Digital Signature #
AC3C984D86FA1BB0
