
Midmark M11 - v1.0.0
Total Cycles: 2292

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2022 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 197.3 F 2.6
4:00 202.4 F 8.0
6:00 228.8 F 9.9
8:00 232.4 F 9.2
10:00 241.4 F 9.8
12:00 243.0 F 12.4
14:00 241.3 F 11.0
16:00 254.6 F 17.7
18:00 265.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.1 F 29.0
4:00 272.5 F 28.8

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 209.8 F 0.0
10:00 226.0 F 0.0
15:00 226.2 F 0.0
20:00 223.6 F 0.0
25:00 223.0 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

09/06/2022 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____