

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4361

Name: RITTER2  
Sterilizer ID: V2313228  
02/19/2025 07:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.5 F 0.0  
2:00 162.5 F 2.3  
4:00 172.2 F 6.6  
6:00 195.9 F 11.3  
8:00 213.5 F 4.9  
10:00 226.3 F 9.8  
12:00 233.4 F 9.3  
14:00 239.9 F 12.1  
16:00 239.5 F 10.8  
18:00 252.3 F 17.4  
20:00 263.8 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 271.7 F 30.0  
2:00 271.9 F 29.6  
3:00 271.2 F 29.5  
4:00 272.2 F 30.0  
  
Min 271.0 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 222.1 F 0.0  
10:00 231.4 F 0.0  
15:00 227.0 F 0.0  
20:00 225.9 F 0.0  
25:00 225.3 F 0.0  
30:00 218.5 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:55  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:00:58  
CYCLE COMPLETE

02/19/2025 08:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_