

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1446

Name: RITTER2  
Sterilizer ID: V2313228  
09/02/2022 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
2:00 201.0 F 2.0  
4:00 199.8 F 6.4  
6:00 217.1 F 11.9  
8:00 226.4 F 7.5  
10:00 238.0 F 13.1  
12:00 239.5 F 11.2  
14:00 238.7 F 10.1  
16:00 252.6 F 17.0  
18:00 264.1 F 23.9  
20:00 271.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.6 F 29.6  
2:00 271.6 F 28.7  
3:00 272.4 F 29.7  
4:00 272.2 F 29.2

Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 218.9 F 0.0  
10:00 237.4 F 0.0  
15:00 233.7 F 0.0  
20:00 222.0 F 0.0  
25:00 222.8 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

09/02/2022 09:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_