
Midmark M11 - v1.0.5
Total Cycles: 3978

Name: RITTER2
Sterilizer ID: V2313228
09/26/2024 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.9 F 0.0
2:00 97.5 F 0.0
4:00 116.3 F 0.1
6:00 140.6 F 0.3
8:00 168.1 F 2.8
10:00 188.7 F 5.9
12:00 207.2 F 9.7
14:00 219.4 F 12.2
16:00 225.6 F 8.0
18:00 235.1 F 12.5
20:00 235.7 F 9.9
22:00 237.3 F 9.6
24:00 246.6 F 14.1
26:00 257.5 F 20.0
28:00 267.5 F 26.3
30:00 273.1 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.8 F 29.9
2:00 271.9 F 29.7
3:00 272.6 F 29.9
4:00 271.8 F 29.7

Min 271.8 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 213.8 F 0.0
10:00 205.7 F 0.0
15:00 224.8 F 0.0
20:00 208.8 F 0.0
25:00 231.6 F 0.0
30:00 233.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:17
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

09/26/2024 03:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____