
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001598
16:43 11/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:51
PRESSURIZING 6:29
STERILIZING 7:54
136.0 C 220kPa 7:54
136.1 C 220kPa 8:24
136.0 C 222kPa 8:54
136.4 C 220kPa 9:24
136.5 C 222kPa 9:54
136.2 C 221kPa 10:24
136.1 C 220kPa 10:54
136.4 C 222kPa 11:24
VENTING 11:24
AIR DRYING 12:28
CYCLE COMPLETE 29:37

Digital Signature #
79AE54727809CCC7
