

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000459  
8:44 02/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:14  
PRESSURIZING 3:23  
STERILIZING 4:30  
136.0 C 227kPa 4:30  
136.3 C 231kPa 5:00  
136.4 C 233kPa 5:30  
136.4 C 232kPa 6:00  
136.2 C 231kPa 6:30  
136.2 C 232kPa 7:00  
136.1 C 230kPa 7:30  
136.4 C 233kPa 8:00  
136.4 C 233kPa 8:00  
VENTING 8:00  
AIR DRYING 8:56  
CYCLE COMPLETE 9:20  
-----

Digital Signature #  
A4073138EDDE5E70  
-----