
Midmark M11 - v1.0.5
Total Cycles: 4236

Name: RITTER2
Sterilizer ID: V2313228
12/20/2024 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 182.4 F 3.1
4:00 196.2 F 8.0
6:00 214.6 F 13.2
8:00 224.9 F 8.1
10:00 238.3 F 13.0
12:00 239.8 F 11.5
14:00 239.9 F 10.7
16:00 253.7 F 17.7
18:00 265.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.7 F 29.5
2:00 272.1 F 29.4
3:00 272.5 F 29.3
4:00 272.2 F 29.5

Min 271.7 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 210.0 F 0.0
10:00 216.8 F 0.0
15:00 218.1 F 0.0
20:00 219.1 F 0.0
25:00 219.7 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

12/20/2024 06:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____