
Midmark M11 - v1.0.0
Total Cycles: 1966

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2022 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 175.0 F 4.1
4:00 194.8 F 8.0
6:00 215.9 F 12.6
8:00 221.3 F 6.1
10:00 233.2 F 10.6
12:00 232.6 F 7.3
14:00 243.6 F 12.8
16:00 241.3 F 11.2
18:00 253.9 F 17.4
20:00 264.8 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.5 F 28.3
3:00 272.3 F 29.2
4:00 272.4 F 28.8

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 226.6 F 0.0
10:00 227.4 F 0.0
15:00 227.9 F 0.0
20:00 226.0 F 0.0
25:00 224.1 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

05/27/2022 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____