

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1306

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2022 12:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.1 F 0.0  
2:00 142.3 F 0.0  
4:00 161.4 F 0.6  
6:00 189.1 F 4.6  
8:00 211.0 F 9.1  
10:00 222.5 F 9.0  
12:00 230.4 F 9.4  
14:00 236.5 F 10.1  
16:00 242.4 F 12.6  
18:00 241.4 F 11.5  
20:00 254.1 F 18.0  
22:00 265.0 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.9 F 29.7  
2:00 272.0 F 28.9  
3:00 272.6 F 29.8  
4:00 272.5 F 29.4

Min 271.6 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
5:00 231.3 F 0.0  
10:00 228.2 F 0.0  
15:00 229.0 F 0.0  
20:00 228.4 F 0.0  
25:00 230.2 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 23:54  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:02:19  
CYCLE COMPLETE

07/15/2022 01:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_