
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004049
8:59 19/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:23
136.3 C 221kPa 4:23
136.3 C 224kPa 4:53
136.1 C 222kPa 5:23
136.3 C 218kPa 5:53
136.3 C 222kPa 6:23
136.2 C 223kPa 6:53
136.3 C 221kPa 7:23
136.4 C 223kPa 7:53
136.4 C 223kPa 7:53
VENTING 7:54
AIR DRYING 9:36
CYCLE COMPLETE 9:42

Digital Signature #
0BACE7F1B6FBC913
