
Midmark M11 - v1.0.5
Total Cycles: 3216

Name: RITTER2
Sterilizer ID: V2313228
12/14/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.1 F 0.0
2:00 172.6 F 2.7
4:00 188.0 F 7.2
6:00 209.0 F 12.6
8:00 220.9 F 7.3
10:00 234.9 F 12.5
12:00 236.4 F 9.8
14:00 237.1 F 8.9
16:00 249.0 F 14.9
18:00 260.8 F 21.6
20:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.4
2:00 271.7 F 28.6
3:00 272.3 F 29.5
4:00 272.4 F 29.1

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 239.9 F 0.0
10:00 211.6 F 0.0
15:00 209.6 F 0.0
20:00 200.2 F 0.0
25:00 194.4 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

12/14/2023 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____