
Midmark M11 - v1.0.0
Total Cycles: 3011

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2023 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 158.8 F 1.3
4:00 168.7 F 3.9
6:00 193.4 F 7.9
8:00 224.8 F 13.1
10:00 226.1 F 7.2
12:00 237.1 F 12.3
14:00 236.4 F 9.4
16:00 239.2 F 9.1
18:00 248.0 F 14.1
20:00 259.9 F 20.8
22:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 271.9 F 28.5
3:00 271.6 F 28.7
4:00 272.5 F 29.1

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 224.0 F 0.0
10:00 234.2 F 0.0
15:00 228.4 F 0.0
20:00 226.6 F 0.0
25:00 225.2 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

03/16/2023 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____