
Midmark M11 - v1.0.0
Total Cycles: 5010

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2024 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.0 F 0.0
2:00 94.2 F 0.0
4:00 119.7 F 0.2
6:00 150.6 F 1.1
8:00 176.3 F 3.3
10:00 190.5 F 6.3
12:00 202.4 F 9.5
14:00 216.2 F 13.1
16:00 220.6 F 6.2
18:00 231.5 F 10.3
20:00 232.8 F 8.5
22:00 241.6 F 11.8
24:00 238.9 F 9.7
26:00 251.3 F 15.7
28:00 261.9 F 21.7
30:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 271.9 F 28.4
2:00 272.5 F 28.5
3:00 271.7 F 27.9
4:00 272.2 F 28.6

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 187.6 F 0.0
10:00 194.7 F 0.0
15:00 194.1 F 0.0
20:00 209.2 F 0.0
25:00 222.9 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:08:55
CYCLE COMPLETE

10/18/2024 02:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____