

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3025

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2023 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.0 F 0.0  
2:00 171.3 F 2.9  
4:00 182.1 F 7.2  
6:00 201.5 F 12.4  
8:00 220.3 F 7.2  
10:00 234.2 F 12.5  
12:00 235.3 F 10.0  
14:00 236.7 F 8.7  
16:00 250.0 F 15.8  
18:00 262.3 F 22.9  
20:00 272.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 273.1 F 29.8  
2:00 271.8 F 29.1  
3:00 271.6 F 29.5  
4:00 272.8 F 29.7  
  
Min 271.2 F 28.9  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 224.6 F 0.0  
10:00 225.6 F 0.0  
15:00 219.8 F 0.0  
20:00 217.4 F 0.0  
25:00 220.7 F 0.0  
30:00 221.2 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:17  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

10/18/2023 11:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_