
Midmark M11 - v1.0.0
Total Cycles: 4067

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2023 03:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.3 F 0.0
2:00 103.1 F 0.9
4:00 116.5 F 3.1
6:00 144.0 F 5.9
8:00 170.7 F 8.9
10:00 187.2 F 12.7
12:00 210.9 F 5.0
14:00 218.6 F 8.9
16:00 230.0 F 13.0
18:00 232.6 F 8.9
20:00 242.5 F 12.4
22:00 242.0 F 11.7
24:00 253.3 F 17.4
26:00 263.4 F 23.4
28:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.9 F 29.5
2:00 271.7 F 28.9
3:00 271.6 F 29.1
4:00 272.9 F 29.4

Min 271.1 F 28.7
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 219.2 F 0.0
10:00 223.6 F 0.0
15:00 218.5 F 0.0
20:00 220.7 F 0.0
25:00 221.9 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 28:19
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:11
CYCLE COMPLETE

11/29/2023 04:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____