

-----  
Midmark M11 - v1.0.5  
Total Cycles: 979

Name: RITTER2  
Sterilizer ID: V2313228  
04/12/2022 11:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.4 F 0.0  
2:00 150.9 F 0.1  
4:00 172.4 F 1.6  
6:00 195.1 F 6.0  
8:00 214.3 F 10.9  
10:00 225.3 F 6.7  
12:00 239.4 F 12.3  
14:00 238.3 F 10.2  
16:00 236.7 F 9.0  
18:00 251.3 F 16.1  
20:00 263.2 F 23.1  
22:00 272.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.5 F 29.2  
2:00 271.9 F 29.1  
3:00 272.4 F 29.1  
4:00 271.9 F 29.2

Min 271.6 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 221.7 F 0.0  
10:00 231.5 F 0.0  
15:00 234.7 F 0.0  
20:00 224.3 F 0.0  
25:00 225.7 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:07  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:19  
CYCLE COMPLETE

04/12/2022 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_